

LABELING DANGEROUS LABORATORY WASTE



MARK OR LABEL THE HAZARD(S).



INCLUDE THE WORDS “**HAZARDOUS WASTE**” OR “**DANGEROUS WASTE.**”



APPLY AN ACCUMULATION START DATE (UNLESS IN A SATELLITE ACCUMULATION AREA).



ONE GALLON OR LESS

- Use text that is appropriately sized for the container.
- Label all containers, including small lab bottles or vials.



MORE THAN ONE GALLON

- Text legible from 25 feet or lettering at least ½” high.
- Hazard pictograms recognizable from 25 feet.

